

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

FEB 12 2010

| | | | |
|--|-------------------------------------|----------------------------------|-------------------------------|
| Supervisor/Operator Performing the Verification Procedure: | | Datamaster cdm S/N <u>130125</u> | |
| Name <u>GARY J. Tellep</u> | | ID# <u>2281</u> | Date <u>2-10-10</u> |
| A | Agency <u>AST</u> | Phone # <u>451-5100</u> | |
| Instrument Location <u>1979 Peger Rd.</u> | | | |
| Alco S/N <u>X301536</u> | | Target Value <u>-078</u> | High Pressure <u>1100 psi</u> |
| B | Alco Test Value Average <u>-073</u> | <u>-074</u> | |
| | 1 st Alco | 2 nd Alco | |
| Signature <u>Gary Tellep</u> | | | |
| (OVER) | | | |

(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz

Scientific Director

State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of February, 2010.

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



State of Arizona
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

130125

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

DIAGNOSTIC CHECK

FOR TEST

| | |
|----------|-------|
| 076 | 23:18 |
| 000 | 23:19 |
| VERIFIED | 23:19 |
| 073 | 23:20 |
| 000 | 23:20 |
| 200 | 23:21 |
| 000 | 23:22 |
| 074 | 23:22 |
| 000 | 23:23 |

— DIAGNOSTIC CHECK —

[illegible]

```
"#<X<>+-.-0123456789!@=>?ABCEFGH  
HIJKLMNOPQRSTUVWXYZ[\\]^_`abcde fghijklmno  
pqrstuvwxyz{|}~!"#$%&'()*+,-./:;<
```